

บริษัท เทสท์ เทคโนโลยี จำกัด TEST TECH CO., LTD.

30, 32 ซอยพระรามที่ 2 ซอย 63 ถนนพระรามที่ 2 แขวงสามยุค เขตบางขุนเทียน กรุงเทพฯ 10150

30, 32 RAMA II Soi 63, Rama II Rd., Samaedam, Bangkhunthian, Bangkok 10150

Tel. 0 – 2893-4211-17 Fax : 0 – 2893-4218

30th
Anniversary



ลำดับ	Parameter	วิธีวิเคราะห์
1	Standard plate Count	Pour Plate
2	Calcium	EDTA Titrimetric Method
3	Magnesium	EDTA Titrimetric, Calculation Method
4	Sodium	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
5	Potassium	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
6	Fluoride	Ion – Selective Electrode Methods
7	Total Chlorine	DPD Colorimetric Method
8	Chloramines	DPD Colorimetric Method
9	Nitrate	Cadmium Reduction Method
10	Sulfate	Turbidimetric Method
11	Copper	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
12	Barium	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
13	Zinc	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
14	Aluminum	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
15	Arsenic	Atomic Absorption Spectrometry (AAS), Hydride Generation / Inductively Coupled Plasma (ICP)
16	Lead	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
17	Silver	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
18	Cadmium	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
19	Chromium	Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)
20	Selenium	Atomic Absorption Spectrometry (AAS), Hydride Generation/ Inductively Coupled Plasma (ICP)
21	Mercury	Cold Vapor, Atomic Absorption Spectrometry (AAS) / Inductively Coupled Plasma (ICP)